

~~NOV 12 2010~~ 11/13/10


# VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

NOV 15 2010


Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130569 ✓
Name	Mac Burrell	ID #	4756 ✓
		Date	11/9/2010 ✓
Agency	Cordova Police Department	Phone #	424-6100
Instrument Location	Cordova Police Department		
Alco S/N	65859 ✓	Target Value	0.081 ✓
		High Pressure	1200
Alco Test Value Average	0.081 ✓		0.081 ✓
	1st Alco		2nd Alco
Signature			
(OVER)			


BMB 12/16/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of DEC, 2010.

  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

DataMaster cdm S/N

130569

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130569

NOVEMBER 09, 2010

OPERATOR'S NAME:

BURRELL/M

OPERATOR'S NUMBER: 4756

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI:

C

O.L. #: C

DEPT/AGENCY: CDVI

CASE/REPORT: 110910

TEST TYPE: V

ALCO TARGET VALUE: .061

ALCO S/N: 65859

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.85 in

ALCO TARGET .000 07:21

BLANK TEST .000 07:22

INTERNAL STANDARD VERIFIED 07:22

ALCO TV 29.85 in .001 07:23

BLANK TEST .000 07:23

SUBJECT SAMPLE .000 07:24

BLANK TEST .000 07:25

ALCO TV 29.86 in .001 07:25

BLANK TEST .000 07:26

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130569

NOVEMBER 09, 2010

TIME 07:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 43c

BAROMETER: 29.86 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~\*